ACCUD

ACCURATE MEASUREMENT




See full product range at: accud.nz

Back Plunger Type Dial Test Indicator

Series 264





Features

Jeweled bearing
Carbide contact point
Two measuring directions
Anti magnetic body
Supplied with two clamps: diameter Ø4mm and Ø8mm
Optional accessory: styli
Graduation: 0.01mm

Specifications

Part No	Range	Accuracy	Hysteresis	Scale
AC-264-008-11	0.8mm	13µm	3µm	0-40-0

© 2026 M Gamer Pty Ltd