ACCUD


ACCURATE MEASUREMENT



See full product range at: accud.nz

Dial Indicator Tester

Series 499



Features

To calibrate dial indicators, test indicators and dial bore gauges
Dimension: 270x210x205mm
Weight: 11kg

Specifications

Part No	Range
AC-499-025-01	Micrometer head Range: 0.25mm Sampling Interval: 0.1mm Accuracy: 4μm(0-25mm), 3μm(any section 10mm), 2μm(any section 1mm) Hysteresis: 1μm Force Gauge Range: 0-1.6N Graduation: 0.05N Accuracy: ±5% Stage Diameter: 0-1.6N Flatness: 1μm Roughness: 0.05μm Half round gauge block (included) Dimension: 10mm Parallelism: 1μm

© 2026 M Gamer Pty Ltd